



SYSTEME SDC POUR HANCHE ET CONDYLE

DHS AND DCS SYSTEM

PLAQUE SDC 95° POUR CONDYLE

DCS PLATE 95°



Matériau : inox ISO 5832-1
Material : stainless steel ISO 5832-1

	Trous	Longueur
	Holes	Length
NMMANZH01002	2	44 mm
NMMANZH01003	3	60 mm
NMMANZH01004	4	76 mm
NMMANZH01005	5	92 mm
NMMANZH01006	6	108 mm
NMMANZH01007	7	124 mm
NMMANZH01008	8	140 mm
NMMANZH01009	9	156 mm
NMMANZH01010	10	172 mm
NMMANZH01011	11	188 mm
NMMANZH01012	12	204 mm
NMMANZH01013	13	220 mm
NMMANZH01014	14	236 mm

Indications : Fractures condyliennes

Indications : Condylar fractures

Espacement des trous : 16,0 mm

Hole spacing : 16.0 mm

Vis préconisées : Vis corticales ø 4,5 mm

Vis spongieuses ø 6,5 mm

Recommended screws : Cortical screws 4.5 mm dia.

Cancellous screws 6.5 mm dia.

